## MILITARY STANDARD-SAMPLING IMPACT RESISTANCE

Procedures and Tables for Inspection by Attributes (MIL-STD-105)

st Records for Inspection Certificate No

Date.

Invoice Number '

Table I-II Sample Size and Acceptance Rejection Criteria

~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~											
NORMAL & TIGHTENED								Ŧ	REDUCE		
BATCH OR LOT SIZE											
SAMPLE		NORMAL		, TI	TIGHTENED		SAMPLE REDUC				
		5	Size	Ac	Re A	rc	Re	Size A	c Re		
				_	_		_				
51	to	90	13	2	3	1	2	5	1 * 3		
91	to	150	20	3	4	2	3	8.	1 * 4		
151	to	280	32	5	6	. 3	4	13	2 * 5		
281	to	500	50	7	8	5	6	20	3 * 6		
501	to	1,200	80	10	11	8	9	32	5*8		
1,201	to	3,200	125	14	15	12	13	50	7 * 10		
3,021	to	10,000	200	21	22	18	19	80	10 * 13		
10,001	to	35,000	315	21	22	18	19	125	10 * 13	}	
35,001	to	150,000	500	21	22	18.	19	200	10 * 13	;	
150,00	1 to	500,000	800	21	22	18	19	315	10 * 13		
500,00	l and	1 over	1,250	21	22	18	19	500	10 * 13		

Ac - Acceptance number

Re = Rejection number

IMPACT RESISTANT TEST RECORDS for the (Company Name)

Sampling Plan NORMAL (Normal & tightened or Reduced)

LENS TYPE QUANTITY SAMPLE SIZE UNBROKEN BROKEN RESULT OF TEST (pcs) (pcs) (pcs) (pcs) Accepted Rejected

Glass Acrylic Polarized Sun sensor , CR 39

<sup>\* =</sup> If the acceptance number has been exceed but the rejection number has not been reached, accept the lot, but reinstate normal inspection.

JUN-10-2009 13:21 WIDER 17182446728 P.002

## INSPECTION CERTIFICATE NO.' (DROP BALL TEST)

## TO WHOM IT MAY CONCERN:

+This is to certify that the following sunglasses (or lenses) are Impact resistant and have passed our for Impact-resistant on the base of a statistically significant sampling pursuant to Sunglass Association of America Standard LT-1 (July 12, 1971) and 21 CFR 801, 410(Meet Current F.D.A. Rulings). The records of our drop ball test are as our attached sheet:

Exporters;	-	-	
Commodities:			•
Quantities:			
Date of Inspection:			
SHIPPING MARKS:			·
			_

Destination:

T2----

F.D.A. Reg. No:

F.D.A. MEDICAL DEVICE LISTING NO.